




DECLARATION OF CONFORMITY

ACCORDING TO (EU) 2017/745 MEDICAL DEVICE REGULATION

EU Representative

SUNGO Cert GmbH

Harffstr. 47, 40591 Düsseldorf, Germany

SRN: DE-AR-000010869

Conformity Assessment

Conformity Assessment Procedure

Annex II+III of Regulation (EU) 2017/745

Applicable Standards

EN ISO 14971: 2019; EN ISO 15223-1: 2021; EN ISO 20417:2021; EN ISO 10993-1: 2020; EN ISO 10993-5: 2009; EN ISO 10993-10: 2023; EN ISO 10993-23: 2021; EN 60601-2-46:2019; EN 60601-1:2006/A2:2021; EN 60601-1-2:2015+A1:2021; EN 62366-1:2015

Remark

The declaration of conformity is valid in connection with the release technical document CE/MDR-Z12011202-01.

All the supporting documentation is retained at the premises of the manufacturer.

The Declaration of Conformity is exclusively under the sole responsibility of the manufacturer.

Manufacturer

Name: Zhejiang Feiman Medical Technology Co., Ltd

Address: Room 3102, Building 3, No. 3088 Century Avenue, Luoxing Street, Jiashan County, Jiaxing, Zhejiang Province

SRN: CN-MF-000037502

Product Information

Name: Electric Hydraulic Operating Table and Accessories

Model: AT500(A-E), AT600(A-E), AT6800(A-E), AT700(A-E), AT800(A-E), AT900(A-E), AT1100(A-E), FP-(001-200), FP-006-(A-E), FP-026-(A-C), FP-028-(A-F), FP-016-(A-B), FP-066-(A-E), QY900(A-E), ZMQY900(A-E), FP-067(A-E), FM-TW-(001-200), WJ(001-1000)

EMDN: Z12011202

Basic UDI-DI: 697683392EHOT0012C

Classification: Class I, According to Rule 13, Annex VIII, Regulation (EU) 2017/745

Declaration

We herewith declare that the above-mentioned products meet the requirements of Medical Device Regulation (EU) 2017/745 and the applicable standards above.

Signature: *Bai Yinhui* Date: 2023.8.7

Position: GM Place: Jiaxing/China

